



Tim Joppich
Merits Australia
Unit 1/17 -19 Musgrave Ave,
Welland, SA 5007

25th May 2015

Dear Tim,

Reference: Review of the Merits Australia W464 Rollator (4 Wheeled aluminium with low seat, loop brake & 6" wheels, indoor) to 114 kg user weight, our Laboratory reference is 492349.

This letter is to advise you that this Laboratory has reviewed the W464 rollator against testing conducted by this Laboratory on the W467 rollator (Our reference for this testing being # 492351 dated 25th May 2015)

The Merits Australia W464 Rollator is identical in all respects to the W467 rollator that was tested by this laboratory with exception to frame width, lower seat height, reduced user weight and fitment of 6" wheels for indoor use only. As the frame design, construction and materials used are identical to the wider frame of the W467 as tested, it is the determination of this test laboratory that the W463 would comply with the same requirements of ISO 11199-2 as the W467 did when it was tested to the substantially higher user weight.

The W467 however, has not been granted full compliance due to the following:

The test sample was not supplied with any labelling / marking addressing the requirements of the following clauses:

- 6.2 (f) Month & year of manufacture
- 6.2 (h) Max width of the rollator
- 6.2 (i) Indoor or outdoor use
- 6.3 Maximum rollator height
- 6.3 Minimum rollator height
- 6.3 Maintenance and cleaning instructions
- 6.3 Instructions for assembly and adjustments
- 6.3 Warning about safe distance between moving and stationary parts
- 6.3 Maximum safe working load for load carrying accessories

Until evidence of the above labelling and marking requirements have been met with the W467, this Test Laboratory is unable to grant compliance to the W464.

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Digitally signed by Wayne Wurfel
DN: cn=Wayne Wurfel, o=Test Laboratory,
ou=NovitaTech, email=wayne.wurfel@novita.
org.au, c=AU
Reason: I attest to the accuracy and integrity
of this document
Date: 2015.05.26 13:53:54 +09'30'

Wayne Wurfel
Senior Test Technician, Test Laboratory

NovitaTech (a division of Novita Children's Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

